



CERTIFICATE

Conformity Assessment for Analysis Equipment

Certificate No.: AK 28108998

Name and address of the Holder : Fer Strumenti S.p.A.
Via Ripamonti, 58
20038 Seregno (MI)
ITALY

Type of equipment : Combustion Product Analyser

Identification : Model : ENOX II

Description : Continuous Analyser for the measurement of :
CO measure range: 0 ... 50 ppm
NO measure range: 0 ... 200 ppm

Reference specification : DLGS 3 Aprile 2006, n. 152 parte V, titoli I e II

Report No.: 28108998

We hereby certify the conformity of the a.m. equipment which complies with the relevant requirements of the reference specification.

During the periodical monitoring, compliance with the specification has been demonstrated. The monitoring will be performed according to the holder's calibration procedure

Milan, 2008-09-18



TÜV CERT - Certification Body for
Pressure Equipment

J. Morenc
Dipl.-Ing. J. Morenc

Notified Body, ID-No. 0035

TÜV Rheinland Industrie Service GmbH
Am Grauen Stein
D-51105 Cologne

TÜV Rheinland Italia
I - 24030 Medolago (BG)
Tel. : +39 035 49.33.147
e-mail: bergamo@it.tuv.com

This Certificate does not imply assessment of the equipment manufacturing and does not permit the use of a TÜV Rheinland mark of conformity